

BRING INNOVATION TO MORE PATIENTS

LENS PARAMETERS

MATERIAL:	samfilcon A
LENS MATERIAL TECHNOLOGY:	MoistureSeal® technology
WATER CONTENT:	46%
OXYGEN PERMEABILITY:	114 Dk
LENS DESIGN TECHNOLOGY:	3-Zone Progressive™ design, OpticAlign™ design
BASE CURVE:	8.6 mm
DIAMETER:	14.5 mm
SPHERICAL POWERS:	+4.00D to -6.00D in 0.25D steps
CYLINDER POWERS:	-0.75D, -1.25D, -1.75D
AXES:	10° to 180° in 10° steps
ADD POWERS:	Low: +0.75D to +1.50D High: +1.75D to +2.50D
VISIBILITY TINT:	Light blue
WEAR SCHEDULE:	Up to 7 days extended wear and daily wear
REPLACEMENT SCHEDULE:	Monthly

Always direct the patient to read and follow the label to use the product properly. For complete product information including approved indication, restriction, cautions and warnings, please refer to product insert.

Bausch & Lomb Canada, Vaughan, Ontario, L4K 4B4
®/™ are trademarks of Bausch & Lomb Incorporated or its affiliates.
© Bausch Health, Canada Inc. TP16570

BAUSCH + LOMB
See better. Live better.

B A U S C H + L O M B U L T R A®

MULTIFOCAL for ASTIGMATISM

FITTING GUIDE

STEP 1: Update spectacle refraction and Add power

STEP 2: Select **toric lens** power
(adjusted for vertex distance if necessary)

STEP 3: Select **Add power** following guidance below:

ADD SELECTION:

SPECTACLE ADD	BOTH EYES
+0.75D to +1.50D	Low Add
+1.75D to +2.50D	High Add

See package insert for more information.

EVALUATE THE LENS FOR SUCCESS

- Allow trial lenses to equilibrate for at least 10 minutes before assessing fit and vision
- Confirm axis orientation
- Evaluate distance and near vision binocularly in normal room illumination
- If vision at distance and near are satisfactory, dispense lenses and schedule follow-up exam within 1-2 weeks



REFINE IF NEEDED

- Confirm axis orientation
- Determine eye dominance
- Follow guidance below



IF PATIENT IS WEARING:

NEAR VISION				DISTANCE VISION		
TWO LOW ADDS	DOMINANT EYE		NON-DOMINANT EYE	DOMINANT EYE		NON-DOMINANT EYE
	Initial Lens	Low Add	Low Add	Initial Lens	Low Add	Low Add
	Refinement 1	Low Add	High Add	Refinement 1	Bausch + Lomb ULTRA® for Astigmatism	Low Add
	Refinement 2: If vision is still unsatisfactory, make small changes by adding +0.25D at a time to non-dominant eye (wearing High Add lens) using hand-held lenses, and continue evaluating vision binocularly in normal room illumination. Adjust contact lens power when vision is satisfactory.			Refinement 2: If vision is still unsatisfactory, make small changes by adding -0.25D at a time to dominant eye (wearing Bausch + Lomb ULTRA® for Astigmatism) using hand-held lenses, and continue evaluating vision binocularly in normal room illumination. Adjust contact lens power when vision is satisfactory.		
TWO HIGH ADDS	DOMINANT EYE		NON-DOMINANT EYE	DOMINANT EYE		NON-DOMINANT EYE
	Initial Lens	High Add	High Add	Initial Lens	High Add	High Add
	Refinement 1	High Add	Add +0.25D to the non-dominant eye	Refinement 1	Low Add	High Add
	Refinement 2: If vision is still unsatisfactory, make small changes by adding +0.25D at a time to non-dominant eye using hand-held lenses, and continue evaluating vision binocularly in normal room illumination. Adjust contact lens power when vision is satisfactory.			Refinement 2: If vision is still unsatisfactory, make small changes by adding -0.25D at a time to dominant eye (wearing Low Add lens) using hand-held lenses, and continue evaluating vision binocularly in normal room illumination. Adjust contact lens power when vision is satisfactory.		